



भारतीय आयुर्विज्ञान अनुसंधान परिषद
स्वास्थ्य अनुसंधान विभाग, स्वास्थ्य और परिवार
कल्याण मंत्रालय, भारत सरकार

Indian Council of Medical Research
Department of Health Research, Ministry of Health
and Family Welfare, Government of India

Dated: 25.07.2023

Corrigendum

Revision of Specifications of Fully automated nitrogen analyzer with digester and accessories

Existing Specifications	Revised Specifications
Fully automated (distillation, colorimetric/pH titration, calculation)	Fully Automated Distillation, inbuilt colorimetric/pH titration and calculation
The distillation unit should equipped with sensors for safety door, tube platform, distillate temperature as well as level and overpressure sensors for the variable output steam generator	The distillation unit should equipped with safety door, tube platform, variable out steam generator and sensor for distillate temperature as well as level.
Steam generator must be made up of glass and should be clearly visible/transparent to check from outside	Steam generator must be glass or metal

--Sd--

Dr. Joy Kumar Chakma
Head, CPC