



icmr
INDIAN COUNCIL OF
MEDICAL RESEARCH
Serving the nation since 1911

NIV
NATIONAL INSTITUTE
OF VIROLOGY

आई सी एम आर - राष्ट्रीय विषाणु विज्ञान संस्थान

स्वास्थ्य अनुसंधान विभाग
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार

ICMR - NATIONAL INSTITUTE OF VIROLOGY

Department of Health Research
Ministry of Health & Family Welfare, Govt. of India

No. LTE PP-498/NIV-Purchase/2024-25

Dt. 30 Jan, 2025

QUOTATION ENQUIRY
(TWO BID)

Sealed quotations are invited on behalf of the Director, ICMR-National Institute of Virology, Pune for "HulaMixer". Interested vendors should submit their quotation along with their company/personal profile and details give below:

Sl. No	Specifications & Details	Qty. Required	Remarks
1	HulaMixer	1	Specifications Attached

The quotation super scribed as "PP no. 498- Hula Mixer" addressed to the Director, ICMR-National Institute of Virology, 20-A, Dr. Ambedkar Road, P.B. No.11, PUNE -411 001. (Attention: Administrative officer - Stores), should be either dropped at the Stores Department at NIV Pune or sent through by Speed Post/Courier latest by 19th Feb, 2025 till 5.30 p.m.

OTHER TERMS & CONDIITONS

1. This is an enquiry and must not be treated as an order.
2. Product details/Brochure along with Technical Bid.
3. Kindly mention Warranty + Comprehensive Warranty details in Technical Bid properly, without mention required information the quotation will be rejected.
4. The NIV, Pune office reserves the right to accept or reject any or all quotations without assigning any reasons.
5. Any decision taken by the Director, NIV Pune at any point of time in connection with this process shall be final and conclusive and no claim or dispute from any bidder in that regard shall be entertained.
6. The bidders are requested to submit Technical Bid and Financial Bid in separate envelops: The bidders submitting in one envelope, will straightaway be rejected.
7. No advance payment will be made. The payment will be made on receipt of services availed and satisfactory report of the end users after installation.
8. NIV, Pune will not be responsible for any delay for receipt of quotation.
9. Requirement of HulaMixer will either increase or decrease final requirement will be mentioned in the P.O order issued to the firm.


Administrative Officer

Detailed Specifications for Sample Mixer Compatible with NGS Library Preparation

PP No: 498/2024-25

1. The equipment should have an adjustable speed range from 1-100 rpm, with orbital rotation, reciprocal rotation and vibration to achieve proper mixing of samples, with rotating angles from 1-90°
2. Minimum of two different platforms must be provided to match tubes of varying volumes ranging from 0.5-50 mL.
3. The equipment should have a wide range of working temperatures [4-40°C necessary]
4. The equipment should be supplied with necessary cables and connectors, compatible with Indian standards, including any necessary adapters. V/A: 120/0.5
5. The equipment should be equipped with an onboard timer, with minimum range of 1-250 seconds
6. The equipment should be compatible with magnetic bead based nucleic acid purification kits.